



Running Contract Details	
Equipment Name	LCD Vision monitor
Running Contract Valid Till	30-12-2020
Tender Ref No	KMSCL/EP/T296/645/2018(R)
Tendered Quantity	50
Supplier Name	M/s Suraj Hi-Tech Pvt Ltd
GST No	32AAACS5018H1ZR
Installation & Delivery Period	4 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
38/2499 - Plappalli House 2nd Floor Gandhi Nagar Kaloor PO Cochin-682017	Contact Person	John Sam
	Phone	
	Mobile No	9995735281
	Email	surajhitechpvtltd@vsnl.net

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	LCD Vision monitor <i>Model & Make : VC-012 / SURAJ HI-TECH PVT LTD.</i>	35840 Incl.GST :12%	2643.2	38483.2
		35840	2643.2	38483.2

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 rd Year	4 th Year	5 th Year
LCD Vision monitor			
Labour	729.00	729.00	729.00
Comprehensive	1,950.00	2,200.00	2,450.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :LCD Vision monitor

1. Should have LCD monitor of 18.5 inches which is remote operated
2. Should have the following charts: Snellen charts, Log Mar chart, Numeric, Alphabets, Dot chart, Contrast sensitivity, Pediatric chart, Masking test, ETDRS contrast Isihara colour test, Cartoon charts, Red Green or duochrome, Peripheral acuity.
3. Should have an available distance of 6 feet, 8 feet, 12 feet, 15 feet and 20 feet
4. Should have eye Educational charts about Lasik, Cataract, Retina, Cornea, Contact lens
5. Should operate from 200 to 240v, 50/60Hz power supply
6. Should have certificate from a competent authority CE/FDA(US)/STQC C3 certificate/ STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid.