



Running Contract Details	
Equipment Name	Laparoscope Set For Sterilization
Running Contract Valid Till	17-06-2021
Tender Ref No	KMSCL/EP/T313/66C/2019
Tendered Quantity	150
Supplier Name	M/s Frontline Electro Medical Ltd
GST No	07AAACF1522F1Z8
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Surajmal Vihar - 110092	Contact Person	Rajesh Kumar Sharma
	Phone	
	Mobile No	9313033505
	Email	rks@feml.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Laparoscope Set For Sterilization Telescope 5mm, 30 degree <i>Model & Make : 8935442 / Richard Wolf</i>	129920 Incl.GST :12%	9581.6	139501.6
2	Ring Applicator 6 to 7 mm	34720 Incl.GST :12%	2560.6	37280.6
3	Trocar & Cannula 7mm	23520 Incl.GST :12%	1734.6	25254.6
4	Trocar & Cannula 5 to 6mm	23520 Incl.GST :12%	1734.6	25254.6
5	Verees pneumoperitoneum needle 8 cm	4480 Incl.GST :12%	330.4	4810.4
6	Verees pneumoperitoneum needle 10 cm	4480 Incl.GST :12%	330.4	4810.4
7	Verees pneumoperitoneum needle 12 cm	3360 Incl.GST :12%	247.8	3607.8

Item-wise Price Details			
	224000	16520	240520

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Laparoscope Set For Sterilization							
Labour	3,000.00	3,000.00	3,000.00	3,000.00	3,000.00	3,000.00	3,000.00
Comprehensive	9,000.00	9,000.00	9,000.00	9,000.00	9,000.00	9,000.00	9,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Laparoscope Set For Sterilization

I. The following items to be supplied as one kit for lap sterilization

1. Telescope 30 degree 5mm

a. Should be a rigid telescope with 5mm diameter

b. Should have a minimum working length of 30 cm.

c. Should be supplied with 1 telescope with viewing direction of and 30°.

d. Should have incorporated fiber optic light transmission.

e. Should have scratch resistant tip.

f. Should be autoclavable

2. Ring Applicator 6 to 7mm

a. Ring applicator, may be fitted with 2 silastic rings. For use with trocar size 6mm

3. Trocar & Cannula

- a. Trocar, Size 7 mm consisting of : Trocar only, with pyramidal tip, cannula without valve, with insufflations stop cock, length 8 cm or more , Multifunctional Valve - 4 Nos
- b. Trocar, Size 5 to 6mm consisting of : Trocar only, with pyramidal tip, cannula without valve, with insufflations stop cock, length 8 cm or more , Multifunctional Valve - 4 Nos

4. Verees pneumoperitoneum needle – Small, medium and Large

- a. Verees pneumoperitoneum needle with spring action blunt stylet, length should be atleast 8 cm, 10 cm and 12 cm, , should have a luer -lock connector for insufflations.

II. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US). Copy of the certificate/ test report shall be produced along with the technical bid