



Running Contract Details	
Equipment Name	Micro Centrifuge for SPH Lab
Running Contract Valid Till	21-12-2020
Tender Ref No	KMSCL/EP/T288/1112/2018(R)
Tendered Quantity	10
Supplier Name	M/s Eppendorf India Limited
GST No	33AABCE1667R1ZA
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Plot NO.18-19-20 (part) Ambit Park Rd Sidco Industrial Estate (South Phase) Ambattur Chennai - 600058	Contact Person	Ganeshprabhu
	Phone	
	Mobile No	9387301010
	Email	ganeshprabhu.s@eppendorf.co.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Micro Centrifuge for SPH Lab</b> <i>Model &amp; Make : 5425 / eppendorf india ltd</i>	218300 Incl.GST :18%	15281	233581
		<b>218300</b>	<b>15281</b>	<b>233581</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Micro Centrifuge for SPH Lab</b>							
Labour	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00	15,000.00
Comprehen sive	45,000.00	45,000.00	45,000.00	45,000.00	45,000.00	45,000.00	45,000.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Micro Centrifuge for SPH Lab**

1. The equipment should be a bench top high speed centrifuge.
2. Should be able to reach RCF of 20,000g.
3. Should be microprocessor controlled with multiple user definable programmes.
4. Should be able to program temperature, RPM, RCF, acceleration and deceleration.
5. The chamber should be stainless steel.
6. Safety features such as lid drop protection and automatic lid lock should be available.
7. Noise level should be below 60dB.
8. Should be supplied with a fixed angle rotor (24 x 1.5/ 2ml) with RCF 20,000 g or above and a PCR strip rotor.
9. Should work with 220v – 230v, 50Hz AC supply