

Running Contract Details	
Equipment Name	Microkeratome with DSAEK Set
Running Contract Valid Till	17-01-2021
Tender Ref No	KMSCL/EP/T288/1245/2018(R)
Tendered Quantity	11
Supplier Name	M/s My Healthskape Medicals Private Limited
GST No	07AADCM4683K1Z3
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Sharma Industrial Estate Walbhat Road Goregoan East Mumbai 400063	Contact Person	Priyansh vora
	Phone	02226862626
	Mobile No	9820444396
	Email	tenders@myhealthskape.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Microkeratome with DSAEK Set <i>Model & Make : MORIA Evo. 3E with DSAEK SET / MORIA S. A</i>	3091200 Incl.GST :12%	227976	3319176
		3091200	227976	3319176

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Microkeratome with DSAEK Set							
Labour	33,750.00	33,750.00	33,750.00	33,750.00	33,750.00	33,750.00	33,750.00
Comprehensive	3,50,000.00	3,50,000.00	3,50,000.00	3,50,000.00	3,50,000.00	3,50,000.00	3,50,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Microkeratome with DSAEK Set

1. Should have automated console.
2. Should have integrated battery backup.
3. Should have automatic change-over of suction pump in situation of any pump failure.
4. Should have disposable head with blade for cutting donor tissue.
5. Should have blade oscillation rate of 15000 rpm.
6. Should have slow vacuum release.
7. Should have facility for connecting wide range of pre-calibrated single use heads of 300,350,400 and 450 microns.
8. Should provide storage and sterilisation box.
9. Should provide autoclaveable and re-useable motor, motor tubing and artificial chamber.
10. Should be able to create thin grafts of less than 300microns.
11. Should supply the standard accessories which includes the following:
 - a. Artificial chamber base-1
 - b. Artificial chamber cover-1
 - c. Moist Chamber-1
 - d. Turbine-1
 - e. Tubings for turbine-1
 - f. Nitrogen Gas Hose-1
 - g. Disposable heads
12. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.