



Running Contract Details	
Equipment Name	Microscope with BF & DF
Running Contract Valid Till	18-07-2020
Tender Ref No	KMSCL/EP/T271/1110/2017(R)
Tendered Quantity	10
Supplier Name	M/s Scientific Enterprises
GST No	32AAIFS7780D1Z4
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	5 Years

Supplier`s Details		
Address	Contact Details	
PB No. 1951 Vytila Kochi â€“ 682019	Contact Person	Deepa
	Phone	
	Mobile No	9447717286
	Email	seqcochin@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Microscope with BF & DF <i>Model & Make : Axio Lab ,L / M/s Carl Zeiss Microimaging GmbH - Germany</i>	554600 Incl.GST :18%	38822	593422
		554600	38822	593422

Annual / Comprehensive Maintenance Charges (Exl.Tax)					
Rate	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Microscope with BF & DF					
Labour	11,210.00	11,210.00	11,210.00	11,210.00	11,210.00
Comprehens ive	27,730.00	27,730.00	27,730.00	27,730.00	27,730.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Microscope with BF & DF

Equipment Name: Microscope with BF & DF

1. The microscope should have provision for BF and dark field.
2. Should have infinity corrected optics with anti-fungal treatment.
3. Should have built in long lasting led illumination.
4. Should have wide field plan corrected 10X eyepiece of minimum 20mm FOV with diopter correction.
5. Should have 4x, 10x, 20x, 40x and 100x full plan corrected objectives. Dark field with good contrast should be available up to 40x magnification without the need to use immersion oil.
6. All objectives should be parfocal and perfectly centered.
7. Slide stage should have ball bearing specimen holder and should have right handed coaxial X and Y movement with travelling range of 75 x 25 mm.
8. Should have an inclined trinocular tube with 30 deg swiveling eyepiece tube.
9. The condenser should have centering controls.
10. Should have a variable field diaphragm.
11. If having separate BF and DF condensers, a turret mount for easy interchange of condensers should be provided.
12. Should have provision for upgrading to fluorescence microscopy with LED illumination.
13. Should have provision for attaching a camera and, a compatible adapter for attaching a Nikon DSLR camera should be supplied.
14. Should provide dust cover for stand.
15. Should provide four preventive maintenance visits per year and all spares including lamps and filters shall be replaced free of cost during warranty and CMC period.
16. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US). Copy of the certificate/ test report shall be produced along with the technical bid.