



Running Contract Details	
Equipment Name	Mobile Spot Light LED Model B
Running Contract Valid Till	09-05-2020
Tender Ref No	KMSCL/EP/T259/74B/2017
Tendered Quantity	625
Supplier Name	M/s Sree Gokulam Healthcare Pvt Ltd
GST No	32AAJCS3502K1ZI
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Dhanya Complex Sastha Temple Road Ernakulam North PO Cochin-682018	Contact Person	Anoop
	Phone	
	Mobile No	9995813333
	Email	sales@gokulamhealthcare.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Mobile Spot Light LED Model B <i>Model & Make : Hepta / LID Sotramet</i>	41328 Incl.GST :12%	3047.94	44375.94
		41328	3047.94	44375.94

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Mobile Spot Light LED Model B				
Labour	2,177.00	2,177.00		2,177.00
Comprehensi ve	4,720.00	4,838.00		4,956.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Mobile Spot Light LED Model B

1. Should be movable on wheel atleast 4 wheels
2. Should be able to adjust position.
3. Should use a LED light source.
4. It should have minimum light intensity of 50,000 lux at 30 cm
5. It should have 4300 chromaticity.
6. CRI should be minimum 90.
7. It should have field size of atleast 11 cm at a distance 30cm from the light head
8. It should have on/OFF switch.
9. Color temperature should be 4000K-5000K
10. Stand height should be minimum 36” from the floor and arm length 30”.
11. Arm should be able to adjust Horizontally and vertically
12. The flexible arm should be adjustable easily to focus at different heights , especially deep inside the patient when in lithotomy position
13. It should have provision to fix with either ICU cot / labor cot / table.
14. Powers supply should be SMPS.
15. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission/FDA(US)/STQC CB certificate /STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate /test report shall be produced along with the technical bid