



Running Contract Details	
Equipment Name	Mobile Spot Light - LED
Running Contract Valid Till	07-04-2021
Tender Ref No	KMSCL/EP/T315/74A/2019(R)
Tendered Quantity	2000
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	10 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Mobile Spot Light - LED <i>Model & Make : Medik LED 1000 / Medik Enterprises</i>	55608 Incl.GST :12%	4101.09	59709.09
		55608	4101.09	59709.09

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Mobile Spot Light - LED				
Labour	1,000.00	1,000.00	1,000.00	
Comprehen sive	2,500.00	2,500.00	2,500.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment : Mobile Spot Light - LED

1. Single dome mobile type with shadow reduction technology
2. Diameter of the dome should be more than 150mm
3. Minimum light output should be at least 50,000 lux at 1 metre distance.
4. Minimum 3 nos of LED should be available.
5. It should have 4300 chromaticity.
6. CRI should be minimum 90.
7. Minimum field size should be 160 mm
8. Should be mounted on at least four caster wheels for free movement
9. Color temperature should be between 3000°K and 5000°K
10. Should have separate stand and arm. The minimum height of the stand shall be 38" from the floor and arm length shall be 36".
11. Arm should be adjustable Horizontally and vertically
12. The light should be adjustable easily to focus at different heights.
13. Should have on/ off switch and the light intensity should be adjustable.
14. Input supply – 230Vac, 50Hz
15. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid