

Running Contract Details	
Equipment Name	Modified Tricycle for Stroke Patients
Running Contract Valid Till	20-10-2021
Tender Ref No	KMSCL/EP/T333/1436/2019(R)
Tendered Quantity	100
Supplier Name	M/s FAIRDEAL SURGICALS
GST No	24AAAFF7462M1Z8
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
30 GALAXY COMMERCIAL JAWAHAR ROAD RAJKOT - 360 001	Contact Person	MR.VYOMESH R.DWIVEDI
	Phone	0281-2224063
	Mobile No	9979849966
	Email	fairdeal_surgicals@yahoo.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Modified Tricycle for Stroke Patients <i>Model & Make : Cogy 1608 RP, Harmony Hill International Co,Ltd</i>	162750 Incl.GST :5%	12803	175553
		162750	12803	175553

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Modified Tricycle for Stroke Patients

1. Should be a tricycle with additional castor wheels for preventing the fall down of patients from the cycle.
2. Should be lightweight and easily maneuverable
3. The cycle should have a tiltable arm rest with a cushion
4. Should have a control switch and it should be connected to left and right side as per requirement
5. The cycle should have minimum three wheels and two of them should have brakes.
6. Should have separate castor wheels for preventing the cycle from tilting.
7. Should have tubeless tyres.
8. Should have pedal straps.
9. Should have a turning control switch/ lever through which the turning of cycle should be controlled
10. Wheels should be easy removable
11. The cycle should be able to bear minimum 100 kg patient weight.
12. The size of the cycle should be minimum 115 x 62 x 88 cm or above
13. The front wheels should be minimum 14 inches and the rear wheels should be minimum 8 inches