

Running Contract Details	
Equipment Name	Modular Cold Chamber
Running Contract Valid Till	05-03-2021
Tender Ref No	KMSCL/EP/T292/1279/2018
Tendered Quantity	25
Supplier Name	M/s Labline Instruments
GST No	32AABFL5818N1Z6
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
P B No. 36 Eroor Road Trippunithura-682301	Contact Person	Bini
	Phone	04842776582
	Mobile No	9020304422
	Email	lablinecochin@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Modular Cold Chamber <i>Model & Make : MC-6S / LABLINE INSTRUMENTS, COCHIN</i>	814200 Incl.GST :18%	56994	871194
		814200	56994	871194

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Modular Cold Chamber							
Labour	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00
Comprehensive	15,000.00	15,000.00	15,000.00	15,000.00	20,000.00	20,000.00	20,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Modular Cold Chamber

1. Should be 6 body capacity and refrigeration of each compartment should be individually controlled, individually insulated and temperature can be adjusted with separate digital display. Individual drainage system should be provided for individual chambers.
2. Construction (inner & outer chamber) -SS304 grade stainless steel
3. Insulation : High Grade Poly Urethane
4. Doors : Made of stainless steel grade 304 sheets /magnetic gasket & External lock
5. Trolleys : Made of stainless steel and sliding
6. Compressor: ISI Marked & CFC Free
7. Defrosting Type : Automatic
8. Insulation : High density poly urethane
9. Removable racks.
10. Molded lockable door catch with internal release.
11. End opening doors
12. Temperature 2° to 8° C.
13. Temperature control : Digital Micro processor
14. Temperature display – LED/LCD display
15. Evaporator : Internal Evaporator system
16. Internal drainage : Should be available
17. Refrigeration system : Air cooled hermetically sealed
18. Air circulation : Forced air circulation
19. Internal lighting : Water proof fluorescent ceiling lamps
20. Power supply : 220 volts / 50 Hz
21. Accessories : Stabilizer of suitable capacity

22. High and low temperature alarm system

23. All 6 chambers should be equal sized and the size of chamber is L- 7 ft X B -3 ft X H- 3ft