

Running Contract Details	
Equipment Name	Monocular inclined compound microscope
Running Contract Valid Till	18-06-2021
Tender Ref No	KMSCL/EP/T324/979/2019(R)
Tendered Quantity	224
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	4 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Monocular inclined compound microscope</b> <i>Model &amp; Make : CXL MONO-LED / N. K. Jain instruments pvt. Ltd..</i>	15859.2 Incl.GST :18%	1110.14	16969.34
		<b>15859.2</b>	<b>1110.14</b>	<b>16969.34</b>

### Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Monocular inclined compound microscope**

1. Should have quadruple nose piece and DIN semi plan Achromatic objectives 10x, 40x (SL), 100x OIL (SL)anti- fungus coating
2. Should have LED illumination with rechargeable battery cells to provide battery back up in power failure
3. Input should be 220-240 VAC
4. Should have steady focus controls with convenient location and ball bearing
5. Should have Monocular tube 45 degree inclined and 360 degree rotatable
6. Should have Integrated wide field eyepiece – 5x/6x, 10 x
7. Should have Anti fungal coated optics
8. Should have Quadruple nose piece with ball bearings
9. Mechanical stage of dimension: 135 x 125 mm, with coaxial control with ball bearing. Graduated scale
10. Should have Coaxial coarse and fine focusing system with ball bearing
11. Illumination: LED illumination with in-built rechargeable battery and (5-6) hour's battery backup in case of electricity failure. Life of the LED source should be at least 1500 hours.
12. Should have Stainless steel holder for slide with 75 x 50 mm movement
13. Should have Abbe condenser, NA 1.25 aspherical lens
14. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL/ BIS. Copy of the certificate/ test report shall be produced along with the technical bid