



Running Contract Details	
Equipment Name	Motorized Suction(Dental)
Running Contract Valid Till	25-06-2020
Tender Ref No	KMSCL/EP/T267/1132/2017(R)
Tendered Quantity	100
Supplier Name	M/s Marvel Agencies
GST No	32AJZPA3421F1ZJ
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
17/1682 D1 Thayyil Arcade Opp. Coronation Theatre Kozhikode -673004 Kerala	Contact Person	
	Phone	
	Mobile No	
	Email	marvelagenciesltd@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Motorized Suction(Dental) <i>Model & Make : Globe 018 / Globe Dent</i>	43349.99 Incl.GST :18%	3034.5	46384.49
		43349.99	3034.5	46384.49

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 rd Year	4 th Year	5 th Year

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned

above.

4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Motorized Suction(Dental)

1. Should be motorized suction with auto flush and auto drain system.
2. Should be made up of steel epoxy coated fitted with oil immersed motorized vacuum pump with two graduated suction jars each of 1.5 Litre capacity with synthetic rubber lids.
3. Should have long pressure tubing.
4. Should have vacuum control.
5. Should have automatic overflow cut off device.
6. Should be mounted on rubber wheels.