

Running Contract Details	
Equipment Name	Movement Therapy Machine
Running Contract Valid Till	02-11-2020
Tender Ref No	KMSCL/EP/T282/686/2018(R)
Tendered Quantity	22
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Movement Therapy Machine</b> <i>Model &amp; Make : TIGO-530 / Thera</i>	428960 Incl.GST :12%	31635.8	460595.8
		<b>428960</b>	<b>31635.8</b>	<b>460595.8</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Movement Therapy Machine</b>							
Labour	7,660.00	7,660.00	7,660.00	7,660.00	7,660.00	7,660.00	7,660.00
Comprehensive	22,980.00	22,980.00	22,980.00	22,980.00	22,980.00	22,980.00	22,980.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Movement Therapy Machine**

1. Should have motor assisted, software controlled movement therapy for upper limb and lower limbs
2. Should have leg support with leg guides
3. Should have adjustable pedals with leg insertion aids
4. Screen display concerning duration of training, distance covered and muscle tone during and after training
5. Should stop automatically if muscle cramps occurs
6. Should provide graphic analysis of leg activity during cycling
7. Should have smooth and quiet operation
8. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.