



Running Contract Details	
Equipment Name	Multifunctional Couch
Running Contract Valid Till	17-08-2020
Tender Ref No	KMSCL/EP/T269/36/2018
Tendered Quantity	250
Supplier Name	M/s Kochin Inlab Equipments India Pvt Ltd
GST No	32AADCK3263F1ZX
Installation & Delivery Period	4 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Door #28/2391-D Vihaan - Chilavannur Kadavanthra P.O Kochi-682020.	Contact Person	Manoj
	Phone	
	Mobile No	9447790938
	Email	manoj.inlab@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Multifunctional Couch</b> <i>Model &amp; Make : COMFY COUCH D40 / QUALIMED SYSTEMS, Thiruvananthapuram</i>	137617.5 Incl.GST :18%	9633.23	147250.73
2	<b>Telescopic IV stand for Chemotherapyâ€“ SS 304 grade</b>	2985.4 Incl.GST :18%	208.98	3194.38
3	<b>Accessory Stand on wheels for Blood Donor couchâ€“ IV stand SS 304 grade</b>	4342.4 Incl.GST :18%	303.97	4646.37
		<b>144945.3</b>	<b>10146.17</b>	<b>155091.47</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year
<b>Multifunctional Couch</b>			
<b>Labour</b>	2,065.00	2,272.00	2,499.00

<b>Annual / Comprehensive Maintenance Charges (Exl.Tax)</b>			
<b>Comprehensive</b>	4,130.00	4,543.00	4,997.00

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Multifunctional Couch**

1. Couch should be made of strong and stable platform.
2. Couch should have a head cushion & arm rest on both sides
3. Length should be at least 2000 mm
4. Width should be at least 650 mm without arm rest
5. Height adjustments between 550 mm to 750 mm
6. Should have motorized height adjustments.
7. Arm rest length should be at least 500 mm
8. Arm rest width should be at least 250 mm
9. Arm rest should be vertical adjustable type, able to be moved out for easy entry of patient / donor.
10. Should be able to adjust position of backrest, seat, leg rest, height and Trendelenburg positions. Conversion to bed position by pressing a single button.
11. Should have a lifting capacity of at least 150 kg
12. Castor wheels – Four Numbers, Diameter of minimum 100 mm, lockable.
13. Couch should have extra thick, disinfectant resistant upholstery and the supplier may submit colour options for upholstery
14. Separate rate shall be offered for Telescopic IV stand for Chemotherapy– SS 304 grade should be provided and it should be mountable on the couch and interchangeable to either side of the patient.
15. Separate rate shall be offered for Accessory Stand on wheels for Blood Donor couch– IV stand SS 304 grade should be fitted on the accessory stand, Minimum 3 shelves, the bottom shelf should be able to hold Blood Collection Monitors of various manufacturers. The

stand provided should have lockable castor wheels.

16. Should provide a wired hand held remote control for the operation of couch.

17. SS304 Foot Rest with padded upholstery - adjustable as per height of the patient / Donor.

18. Handle should be provided on the rear side back rest for easy transportation of chairs

19. Should work with input 220 to 240Vac 50 Hz supply.

20. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

21. Minimum dimensions of the Accessory stand for Transfusion Blood Collection Monitor Tray size : 400mm x 280mm  
Sphygmomanometer tray size : 370 mm x 130mm

Accessories Tray Size : 370 mm x 160mm.