

| Running Contract Details | |
|--------------------------------|---------------------------|
| Equipment Name | Myoma morcellator |
| Running Contract Valid Till | 26-04-2020 |
| Tender Ref No | KMSCL/EP/T262/623/2017(R) |
| Tendered Quantity | 10 |
| Supplier Name | M/s J S Surgicals |
| GST No | 32DKOPS2430B1ZG |
| Installation & Delivery Period | 6 Week(s) |
| Up-time / PM vist | 95% & 4 Visits per year |
| Warranty period | 3 Years |

| Supplier`s Details | | |
|--|-----------------|-----------------------|
| Address | Contact Details | |
| TC NO 8/3031/13 DALLYS APARTMENT CHALAKUZHAY ROAD PATTOM (PO) TRIVANDRUM - 695004 | Contact Person | Satheeshkumar S |
| | Phone | 04712443666 |
| | Mobile No | 9995341646,9388227592 |
| | Email | jssurgicals@gmail.com |

| Item-wise Price Details | | | | |
|-------------------------|---|---|------------------------------------|---------------|
| # | Item Details | Unit Rate (Incl.all taxes & charges) | Service Charges (Through KMSCL) | Grand Total |
| 1 | Myoma morcellator <i>Model & Make : Versator Morcellator / Veol Medical Technologies Pvt. Ltd</i> | 476000 Incl.GST :12% | 35105 | 511105 |
| | | 476000 | 35105 | 511105 |

| Annual / Comprehensive Maintenance Charges (Exl.Tax) | | | | | | | |
|--|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|-----------------------|
| Rate | 4 th Year | 5 th Year | 6 th Year | 7 th Year | 8 th Year | 9 th Year | 10 th Year |
| Myoma morcellator | | | | | | | |
| Labour | 23,600.00 | 23,718.00 | 23,837.00 | 23,955.00 | 24,076.00 | 24,196.00 | 24,317.00 |
| Comprehensive | 48,380.00 | 48,622.00 | 48,865.00 | 49,109.00 | 49,355.00 | 49,601.00 | 49,849.00 |

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Myoma morcellator

1. The equipment shall consist of motor system, two pedal switch, connecting cord for use with morcellator.
2. Shall have Motor system, power supply : 230-240 VAC, 50/60 Hz
3. Maintenance of the motor driven components (motors, handles, shaver hand pieces) shall be performed with universal spray
4. Spacer - package of 5 pieces - maximum of height clearance 10mm, for use with instruments up to 16.5mm diameter
5. Morcellator for laparoscope application- Set, diameter 15mm cutter, including
 - a. Hollow shaft motor obturator, with blunt tip, 15mm.
 - b. Trocarbush, askew, 15mm, handle, laparoscopic cutter, laparoscopic, 15mm
 - c. Motor valve 15mm
 - d. Sealing cap for single use, 10 pieces
 - e. Clickline tenaculum Forceps, diameter 15mm for use with Motor System
6. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid