

Running Contract Details	
Equipment Name	Nebuliser Ultrasonic
Running Contract Valid Till	13-12-2020
Tender Ref No	KMSCL/EP/T296/253/2018(R)
Tendered Quantity	100
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	4 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	0 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Nebuliser Ultrasonic <i>Model & Make : NE-U 780 / Omron</i>	34832 Incl.GST :12%	2568.86	37400.86
		34832	2568.86	37400.86

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Nebuliser Ultrasonic

- a. Should be light weight, portable, Compact and easy to use.
- b. Frequency of ultrasonic generator should be greater than 2.5 MHz
- c. Should have 3 speed nebulization rate control (minimum, medium, maximum)
- d. Should have a nebulisation capacity of 0.3 ml/min.
- e. Transducer element should have life of at least 5000 hours
- f. Medication cup capacity should have capacity of maximum 8ml.
- g. Should uses water as ultrasonic conduction medium, no gel is required.
- h. Should provide silent operation.
- i. Should have a built in timer.
- j. Should works on 200-240 VAC / 50 Hz.
- k. Should be provided with a complete nebulisation kit of 10 Nos. including adult and child mask and medication cup – 5 Nos.
- l. Should have safety certificate from a competent authority CE / FDA (US) or any other equivalent standards.