



Running Contract Details	
Equipment Name	Non mydriatic camera
Running Contract Valid Till	18-02-2020
Tender Ref No	KMSCL/EP/T247/994/2017
Tendered Quantity	20
Supplier Name	M/s Appasamy Associates
GST No	33AASCA3660R1Z0
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
# 20 - SBI Officers Colony 1st Street Arumbakkam Chennai-600106	Contact Person	B Radhakrishnan
	Phone	04430101400, 30101401
	Mobile No	09349211170
	Email	info@appasamy.com, appasamy.thrissur@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Non mydriatic camera <i>Model & Make : CR-2 AF / Canon Singapore PTE.Ltd</i>	1008000 Incl.GST :12%	74340	1082340
		1008000	74340	1082340

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Non mydriatic camera							
Labour	31,860.00	42,480.00	53,100.00	63,720.00	74,340.00	84,960.00	95,580.00
Comprehensive	84,960.00	95,580.00	1,06,200.00	1,16,820.00	1,27,440.00	1,38,060.00	1,48,680.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Non mydriatic camera

1. Should have a field angle of 45°.
2. Should have colour photography mode.
3. Should have focusing using split luminous bars coincidence.
4. Should have internal and external fixation target.
5. Should have 5.5" or better LCD /LED monitor.
6. Should have a storage capacity of 1TB hard disk.
7. Light source:

Observation: near- infrared LED

Photography: Xenon flash lamp.

8. Compensation range of examined eye:

- Without compensation: -12D to 12D
- Negative compensation: -32D to -10D
- Positive compensation: 10D to 35D

9. Should have optical adjustment range:

- Forward/backward=40mm
- Leftward/rightward=98mm
- Vertical=27mm

10. Should have chin rest adjustment of 55mm or more.

11. Should operate on mains 220-240Vac, 50 Hz single phase.

12. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid.