



Running Contract Details	
Equipment Name	OT Light shadow-less ceiling mount Single dome
Running Contract Valid Till	22-01-2021
Tender Ref No	KMSCL/EP/T298/90/2018(R)
Tendered Quantity	100
Supplier Name	M/s Sree Gokulam Healthcare Pvt Ltd
GST No	32AAJCS3502K1ZI
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Dhanya Complex Sastha Temple Road Ernakulam North PO Cochin-682018	Contact Person	Anoop
	Phone	
	Mobile No	9995813333
	Email	sales@gokulamhealthcare.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	OT Light shadow-less ceiling mount Single dome <i>Model & Make : Crelite 500 / Shandong Lewin Medical Equipment Co.,Ltd.</i>	120428 Incl.GST :12%	8881.57	129309.57
		120428	8881.57	129309.57

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
OT Light shadow-less ceiling mount Single dome							
Labour	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00
Comprehensive	20,000.00	21,000.00	22,000.00	23,000.00	24,000.00	25,000.00	26,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :OT Light shadow-less ceiling mount Single dome

1. Single dome ceiling type Operation Theatre light shadow-less
2. Light output should be minimum 1,00,000 lux at 1 meter distance
3. Diameter of the dome should be less than 650mm.
4. The dome should have single bulb with standby bulb. The standby bulb can be turned on automatically incase of failure of the primay bulb. Should have indication for bulb failure.
5. Minimum light field size should be 200mm
6. Power supply to the lamp should be from a constant voltage source/SMPS.
7. Colour temperature should be between 4000ËšK to 5000ËšK
8. Should have short circuit protection for the low voltage power supply
9. Should have user friendly lamp replacement
10. Should have heat reduction filters
11. Should have field focus control through sterile handle. Additional sterile handles two numbers should be supplied
12. Five spare bulbs should be supplied.
13. Minimum spring arm stroke of 500mm.
14. Minimum four step intensity control
15. Should have minimum action radius of 1500mm
16. Input supply – 230Vac, 50Hz.
17. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid