

Running Contract Details	
Equipment Name	Operating Loupes
Running Contract Valid Till	07-07-2021
Tender Ref No	KMSCL/EP/T324/1212/2019(R)
Tendered Quantity	25
Supplier Name	M/s Prathosh medico
GST No	32AAIFP5061A1ZR
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
Kousthubham 26/209B Chitrakalalayam Road AKG Road Extn: Edappally Toll Kochi - 682024	Contact Person	Mr. P.J. Jayaprakash
	Phone	
	Mobile No	9447052710
	Email	prathoshmedico@yahoo.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Operating Loupes <i>Model & Make : Hires Plus Loupes Medium / Orasoptic</i>	212800 Incl.GST :12%	15694	228494
		212800	15694	228494

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Operating Loupes

- a. Magnification should be 3.5 times (Range of 3-4)
- b. It should be through the lens and prismatic optical design.
- c. Glass lenses should be fused with Anti scratch and Anti reflection coatings
- d. All Carrier lenses should made with intex poly-carbonate material
- e. It should be added with User Eye prescription.
- f. Field of view and depth should be exactly at 10.2cm
- g. Loupe weight should be less than 80 Grams.
- h. The complete loupe kit should be supplied along with Head strap, carrying case, cleaning cloth, Side shields, protective lens cap, optical screw driver.
- i. Loupe should be patented variable focus feature to adjust the fine tune the user exact working distance (Range of working distance 42-46 cm).
- j. Should have Water Proof & Micro Size Optics
- k. Should have maximum View of operating site with custom working distance.