



Running Contract Details	
Equipment Name	Operating Microscope ENT Model B
Running Contract Valid Till	03-05-2020
Tender Ref No	KMSCL/EP/T262/86A/2017(R)
Tendered Quantity	4
Supplier Name	M/s Carl Zeiss India (Bangalore) Pvt Ltd
GST No	29AADCC6152H1ZM
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
55/2744 4A Santhi Emerald Cheruparambeth Road Kadavanthra Cochin â€™ 682 020. Kerala	Contact Person	Mr. Madhu.P
	Phone	484 44061470
	Mobile No	9446300080
	Email	madhu.poovakkat@zeiss.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Operating Microscope ENT Model B <i>Model & Make : OPMI 1 FC / Carl Zeiss Meditec, AG</i>	2948218.2 Incl.GST :18%	206375.27	3154593.47
		2948218.2	206375.27	3154593.47

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Operating Microscope ENT Model B							
Labour	23,000.00	24,150.00	25,357.00	26,625.00	27,956.00	29,354.00	30,822.00
Comprehen sive	80,000.00	84,000.00	88,200.00	92,610.00	97,240.00	1,02,102.00	1,07,207.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Operating Microscope ENT Model B

1. Manual 5 step magnification changer with apochromatic optics and integrated green filters
2. Objective lenses f=200mm, 300 mm (with manual focusing mechanism and sterilized rubber cap range minimum 13mm) and f=400mm (mount diameter 48mm)
3. Straight binocular tube – f=170mm, interpupillary distance adjustment from 55mm to 75mm
4. Pair of wide field push in eyepieces 10 x /12x with sleeves and magnetic locks, diopter setting from -5D to +5D
5. Integrated manual fine focusing adjusting knob, focusing range should be 40 mm approximately
6. With objective lenses f= 200mm and eyepiece 10 x magnification -3.4x-5.1x-8.5x-13.6x -21.3x- visual field diameter 61.8 - 41.2 --24.7- 15.5 -10.0mm
7. 120 inched coupling
8. Front slide hand grips
9. Coaxial fibre optic illumination system
10. Dual lamp system with quick change over (Not applicable for LED Light Source)
11. Continuously lamp intensity adjustment by control knob near to the surgeon
12. Automatic on/off switch for the lamp
13. Stable and sturdy floor stand on four lockable castors, column (height minimum 1.7m), spring balanced articulated arm, carrier arm power supply unit light guide 2m, power cable 6m
14. Sterilised rubber caps for all knobs and dust cover
15. Stereo co observation attachment
16. HD Camera with HD TV of size 32 inch with C Mount and beam splitter.
17. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid