

Running Contract Details	
Equipment Name	Ophthalmic Operating Microscope
Running Contract Valid Till	26-11-2020
Tender Ref No	KMSCL/EP/T282/87/2018(R)
Tendered Quantity	54
Supplier Name	M/s Towa Sales Corporation
GST No	07AAAFT5502A1ZU
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
# 223 Okhla Industrial Estate New Delhi - 110020	Contact Person	
	Phone	011-26928913
	Mobile No	09810273742
	Email	towasls@md4.vsnl.net.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Ophthalmic Operating Microscope <i>Model & Make : OM-9 / Takagi Seiko Co.Ltd</i>	1016524.32 Incl.GST :12%	74968.67	1091492.99
		1016524.32	74968.67	1091492.99

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Ophthalmic Operating Microscope							
Labour	18,222.00	20,044.00	22,048.00	24,253.00	26,678.00	29,346.00	32,281.00
Comprehensive	36,444.00	40,088.00	44,097.00	48,507.00	53,358.00	58,694.00	64,563.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Ophthalmic Operating Microscope

1. Should have apochromatic and imported optics
2. Should have 5 step magnification
3. Should have working distance of objective lens of focal length $F = 200\text{mm}$
4. Eye piece should be minimum 10x ,12.5x or 16x wide
5. Should have XY coupling
6. Should have tiltable binocular with converging optics.
7. Should have field of view from at least 15mm to 50mm
8. Should have cold light/LED coaxial illumination
9. Should have tools free design for stand-by bulb change over and for failed bulb replacement if required.
10. Should have heat absorbing and UV filters.
11. Should be floor standing type with fiber wheels with brake.
12. Should have a minimum vertical stroke of 400mm
13. Should have rust free stand.
14. Should be operated in 200-240 Vac 50/60 Hz input supply.
15. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid