



Running Contract Details	
Equipment Name	Optical Coherence Tomography Model B
Running Contract Valid Till	18-01-2021
Tender Ref No	KMSCL/EP/T284/607B/2018(R)
Tendered Quantity	10
Supplier Name	M/s Appasamy Associates
GST No	33AASCA3660R1Z0
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
# 20 - SBI Officers Colony 1st Street Arumbakkam Chennai-600106	Contact Person	B Radhakrishnan
	Phone	04430101400, 30101401
	Mobile No	09349211170
	Email	info@appasamy.com, appasamy.thrissur@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Optical Coherence Tomography Model B <i>Model & Make : Appa OCT 2015 / Appasamy Associates</i>	1799952 Incl.GST :12%	132746.46	1932698.46
		1799952	132746.46	1932698.46

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Optical Coherence Tomography Model B							
Labour	32,000.00	32,000.00	48,000.00	48,000.00	64,000.00	64,000.00	80,000.00
Comprehensive	1,12,000.00	1,12,000.00	1,28,000.00	1,28,000.00	1,44,000.00	1,44,000.00	1,60,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Optical Coherence Tomography Model B

1. Methodology - Spectral Domain OCT . Both Anterior segment& Posterior segment
2. Optical Source – Super Luminescent Diode and Infrared for Fundus Imaging
3. Vertical Resolution - 12 to 15 Microns
4. Horizontal Resolution - 5 Microns or better
5. Field of View - 30 degrees or above
6. Scan depth - 2mm into Retinal Tissue
7. In Anterior Segment - Corneal Thickness Measurement
8. Power Supply – 230 Volts AC
9. Internal and External fixation - Provision should be available
10. Scan pattern - Macula 3 D, Disc Analysis
11. Pupil Diameter - 3-3.5mm size

Viewing of Picture and Storing of Picture

1. Computer with required UPS&LCD colour monitor 24"
2. Printer (colour Laser)
3. Motorized Table
4. Storage Facility
5. Online UPS for 1 hour Back up