



Running Contract Details	
Equipment Name	Otoscope
Running Contract Valid Till	27-12-2020
Tender Ref No	KMSCL/EP/T298/94/2018(R)
Tendered Quantity	500
Supplier Name	M/s Trivandrum Surgicals and Lab Supplies
GST No	32AAIFT1010D1Z0
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
T.C.2/1094(7) SREELAKSHMI PAZHAYA ROAD MEDICAL COLLEGE P.O TVM.695015	Contact Person	GEETHA . S
	Phone	0471 - 2555040
	Mobile No	09447340147
	Email	trivandrumsurgicals@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Otoscope <i>Model & Make : K100 / heine</i>	24640 Incl.GST :12%	1817.2	26457.2
		24640	1817.2	26457.2

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Otoscope				
Labour	1,540.00	1,760.00	1,980.00	
Comprehensive	1,540.00	1,540.00	1,540.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Otoscope

1. Should be a C-Cell type or D-cell type Otoscope suitable for usage in outpatient departments of Govt. Hospitals.
2. Should be provided with a halogen/LED light source. In case of halogen light source, 2 Nos of spare bulbs are to be provided.
3. Should be able to detach the otoscope head.
4. Should provide no reflections and obstructions.
5. Should provide the following detachable accessories
 - a. Specula (2.5mm to 5mm, 4 sizes)
6. Should have in built rechargeable battery. Recharge should be possible with direct mains supply
7. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid