



Running Contract Details	
Equipment Name	Oxygen concentrator
Running Contract Valid Till	13-06-2020
Tender Ref No	KMSCL/EP/T264/95/2017
Tendered Quantity	200
Supplier Name	M/s Moditech International
GST No	04ABFFM2524D1ZU
Installation & Delivery Period	4 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
4a Metro Plaza Sco 54-55 Sector 9 D Chandigarh	Contact Person	Mohit Arora
	Phone	
	Mobile No	9216464081
	Email	moditech1@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Oxygen concentrator <i>Model & Make : JAY 5 / LONGFIAN SCITECH CO. LTD CHINA</i>	29904 Incl.GST :12%	2205.42	32109.42
		29904	2205.42	32109.42

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Oxygen concentrator				
Labour	590.00	590.00	590.00	
Comprehensive	2,950.00	2,950.00	2,950.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Oxygen concentrator

1. Oxygen concentrator to provide oxygen from ambient air
2. Oxygen concentration measured at the flow meter by oxygen sensing device (OSD)
3. Sound level <50 dB
4. Superior grade of molecular sieve
5. Maintenance free valve.
6. Oxygen purity, approx: 90%
7. Oxygen output, approx: 0 - 5 LPM
8. Pressure, approx: 8 psi
9. Double outlet or flowsplitter for oxygen Delivery
10. Oxygen tube of 2 m length must be provided with
11. Facility for nebulization with tube & mask
12. With two humidifier bottles and two cabinet filters
13. Unit should function with 200-240Vac, 50/60 Hz input power supply
14. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
15. The equipment shall be supplied with
 - (i) One spare set of tubing
 - (ii) One spare set of internal and external filters (bacterial)
 - (iii) One spare set of fuses
 - (iv) User manual with trouble shooting guidance, in English
 - (v) Technical manual with maintenance and first line technical intervention instructions, in English

(vi) List of priced accessories and priced spare parts