



Running Contract Details	
Equipment Name	PAGE Electrophoresis System
Running Contract Valid Till	24-04-2021
Tender Ref No	KMSCL/EP/T307/1327/2019(R)
Tendered Quantity	15
Supplier Name	M/s Crescent Lab Equipments
GST No	32AAFFC9887N1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
BMRA-136 B Theckineth B.K Menon Road Edappally Cochin -682024 Kerala	Contact Person	
	Phone	
	Mobile No	
	Email	crescentlab07@gmail.com, info@crescentlab.in,

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>PAGE Electrophoresis System</b> <i>Model &amp; Make : 1658003 / BIORAD</i>	116235.9 Incl.GST :18%	8136.51	124372.41
		<b>116235.9</b>	<b>8136.51</b>	<b>124372.41</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year	
<b>PAGE Electrophoresis System</b>				
Labour	12,000.00	13,200.00		14,520.00
Comprehensive	18,000.00	19,800.00		21,780.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :PAGE Electrophoresis System**

1. Mini Protein electrophoresis system (dual).
  - i. Gel size: 8 x7 cm.
  - ii. Spacers of 0.75 & 1.0mm thickness fused with glass plates, sample loading guides.
  - iii. Two gel running capacity.
  - iv. Ridged combs 10 & 15 well.
  - v. Gel casting stand, no tapes or grease.