



Running Contract Details	
Equipment Name	PH Meter for Neuro Science Lab
Running Contract Valid Till	22-02-2021
Tender Ref No	KMSCL/EP/T305/823/2019
Tendered Quantity	20
Supplier Name	M/s Servell Bio Engineers Pvt Ltd
GST No	33AARCS1684A1ZE
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.22/11 Bazaar Road Saidapet Chennai Tamilnadu 600015	Contact Person	R.Sivaraman
	Phone	04424344025
	Mobile No	9094042760
	Email	sales@servellbio.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	PH Meter for Neuro Science Lab <i>Model & Make : LMPH 12 / Labman Sceintific Instruments Pvt. Ltd</i>	36816 Incl.GST :18%	2577.12	39393.12
		36816	2577.12	39393.12

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
PH Meter for Neuro Science Lab							
Labour	1,000.00	1,000.00	1,500.00	1,500.00	1,500.00	2,500.00	2,500.00
Comprehensive	1,500.00	2,500.00	3,500.00	3,500.00	3,500.00	3,500.00	3,500.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :PH Meter for Neuro Science Lab

1. Should be with multi-point push button calibration with auto buffer recognition
2. Dual display of pH or mV with temperature
3. Should provide compatible electrode
4. Should provide compatible USA or NIST buffer standards
5. Should be with built in electrode stand and built-in electrode arm
6. Temperature probe should be included
7. Self-diagnostic messages
8. Stores up to 100 sets of data
9. Should have
10. pH Range : 0.00 to 14.00 pH
11. Resolution/Accuracy : 0.01 pH/ ± 0.01 pH
12. mV Range : ± 1999 mV
13. Resolution/Accuracy : (± 199.9 mV) / 1 mV ($>\pm 199.9$ mV) ± 0.2 (+ 199.9 mV) / ± 2 mV ($>\pm 199.9$ mV)
14. Temperature range : 0.0 to 1000.0 C
15. Resolution/Accuracy : 0.10 0 C(0.10 0 F)/+ 0.30 0 C (0.5 0 F)
16. No.of pH Calibration Points : 5 points
17. Buffer Options : USA; pH 1.68, 4.01, 7.00, 10.01, 12.45 NIST; pH 1.68, 4.01, 6.86, 9.18, 12.45
18. Power Requirements : Included, adapter for 220 VAC,500 mA,9 VDC
19. All accessories needed to make the equipment functional, must be supplied