



Running Contract Details	
Equipment Name	Pacemaker (Temporary) - Dual Chamber
Running Contract Valid Till	22-06-2020
Tender Ref No	KMSCL/EP/T267/303B/2017(R)
Tendered Quantity	20
Supplier Name	M/s Sree Gokulam Healthcare Pvt Ltd
GST No	32AAJCS3502K1ZI
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Dhanya Complex Sastha Temple Road Ernakulam North PO Cochin-682018	Contact Person	Anoop
	Phone	
	Mobile No	9995813333
	Email	sales@gokulamhealthcare.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Pacemaker (Temporary) - Dual Chamber <i>Model & Make : Pace 203 / Osypka Medical</i>	345520 Incl.GST :12%	25482.1	371002.1
		345520	25482.1	371002.1

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Pacemaker (Temporary) - Dual Chamber				
Labour	11,800.00	11,800.00		11,800.00
Comprehens ive	25,960.00	25,960.00		25,960.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Pacemaker (Temporary) - Dual Chamber

1. Should be a Dual Chamber Pacemaker (Temporary) for bradycardia treatment before, during or after a surgery.
2. Stimulation burst and permanent stimulation should be available for high pacing rate.
3. Should be compact & easy-to-operate device, particularly suitable for emergency treatments.
4. Safety features, including automatic lead and battery check.
5. Should have continuous monitoring of the battery status.
6. Should have transparent cover for parameter protection.
7. Should have shock and water-resistant housing.
8. Should have back up pacing during battery change.
9. Should have Modes AOO, AAI, VOO, VII, DDD, DDI/equivalent, DVI
10. Should have pacing rate 30 – 200 ppm.
11. Should have fast pacing (Burst rate) of 80-200 ppm.Should have AV Delay 20 – 300 ms.
12. Should have Pulse Amplitude (A/V) – 0.1 – 17.0.
13. Should have Sensitivity (A/V) 0.4-10.0 mV (A) and 1.0-20.0 mV (V).
14. Should have minimum battery backup > 150 hours.
15. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.