



Running Contract Details	
Equipment Name	Pediatric CPR Manikin
Running Contract Valid Till	20-03-2021
Tender Ref No	KMSCL/EP/T307/1305/2019(R)
Tendered Quantity	263
Supplier Name	M/s Laerdal Medical India Pvt Ltd
GST No	33AABCL8105F1ZX
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.10 1st Floor 1st Street Dr. Subbarayan Nagar Kodambakam Chennai- 600 024.	Contact Person	Balaganesh
	Phone	044-42614773
	Mobile No	9840153367
	Email	balaganesh.mahendrathas@laederal.co.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Pediatric CPR Manikin <i>Model & Make : Little Jr QCPR / Laerdal Medical Norway</i>	153400 Incl.GST :9%	10738	164138
		153400	10738	164138

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Pediatric CPR Manikin

1. The Pediatric Airway Management Trainer along with CPR Manikin should provide the training of CPR with feedback on the performed CPR. It should have an option to accommodate training in Rescue/ extrication or Life Saving First Aid.
2. Airway must open/close by the following procedures when they are performed correctly as taught according to ILCOR guidelines:
 - a) Head tilt
 - b) Chin lift
 - c) Jaw thrust
3. Ventilation of the manikin must be possible through the following procedures:
 - a) Mouth to mouth
 - b) Mouth to nose
 - c) Mask to mouth (both Pocket Mask and Bag-Valve Mask (BVM))
4. The manikin must show a realistic chest rise during ventilation
5. Removal and replacement of lungs and face skin must be easy without use of tools.
6. Ears must be visible as a landmark for correctly aligning the head in order to apply a cervical spine extrication collar. Should have collar bones, sternal notch and belly button.
7. The manikin must be equipped with means of providing audible feedback on performed correct CPR.
8. Should be a realistic life size intubation trainer with a flexible tongue arytenoid cartilage epiglottis vocal cords, trachea, esophageous and simulated lung. The manikin must be equipped with means of providing visual feedback on performed CPR. The head should be capable of full range of tilt and rotation.
9. The manikin should have a completely removable chest cover, compression spring and face mask.
10. The manikin should have a disposable lower airway with an integral one way valve.
11. The individual manikin configuration should come standard with a carry Bag which also serves as a padded training mat.