



Running Contract Details	
Equipment Name	Pharmacy Refrigerator (1000 Litre)
Running Contract Valid Till	23-02-2022
Tender Ref No	KMSCL/EP/T346/922/2019
Tendered Quantity	50
Supplier Name	M/s New India Surgicals and Chemicals
GST No	32AAGFN8445E1ZD
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
30/1212IV Prasanth Buildings Near Medical College Hospital Calicut-673008	Contact Person	
	Phone	
	Mobile No	
	Email	nisc2000@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Pharmacy Refrigerator (1000 Litre) <i>Model & Make : FRC-V-1000 , Cryo Scientific Systems Pvt.Ltd</i>	210040 Incl.GST :18%	14702.8	224742.8
		210040	14702.8	224742.8

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Pharmacy Refrigerator (1000 Litre)				
Labour	15,000.00	15,000.00	15,000.00	
Comprehensive	25,000.00	25,000.00	25,000.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Pharmacy Refrigerator (1000 Litre)

1. Should have upright type unit.
2. Should have a storage Capacity of (1000 –1400) liters.
3. Should have adjustable temperature control range from +2 to +8 degree Celsius.
4. Should have a hermetically sealed type Compressor.
5. Refrigerant System-
 - a) CFC Free Eco-friendly refrigerant R134a/R404a or equivalent.
 - b) The system should have positive, forced, air circulation to maintain temperature uniformity at all shelf levels, with quick recovery +/-1 degree Celsius.
 - c) The system should have sensors for activating automatic defrost cycle to minimize the frost build up.
6. Should have a temperature accuracy of +/-0.3degree.
7. The Inner Cabinet should be made of durable white epoxy painted steel.
8. The Outer Cabinet made up powder coated corrosion resistant sheet.
9. The Door system should have lockable double glass doors for better safety, (self closing hinge doors)
10. External ambient temperature:- should perform in ambient temperature up to +45 degree Celsius.
11. System should have large and clear digital displays for the set/run parameters.
12. Alarms
 - a) System should have audible/visual warnings for over-temperature, under temperature, power failure and low battery.
13. Should have Clear product visibility with dual glasses.
14. Should have adjustable plastic coated wire mesh shelves.
15. Number of shelves- 5-6 Numbers
16. Power input :- 220-240V AC, 50Hz with power cord & plug
17. Automatic Voltage stabilizer with time delay restarts of appropriate ratings meeting ISI Specifications (Input 160-260V and output

220-240V and 50Hz).

18. Should have FDA/ CE/ UL or BIS approved product.
19. The unit should have fitted with castor wheels, front two with brakes.
20. Should have No frost, fan-assisted cooling circulation system.
21. Ultra fast pull down, Auto defrost with self-evaporating drip tray.
22. Should have Rear-to-front hot air anti-condensation.
23. Should have Digital temperature controller display.
24. Should have Full length illumination.
25. Should have Low noise, Green technology & Energy saving (4/5 star rating).