



Running Contract Details	
Equipment Name	Photo therapy unit single surface LED Type
Running Contract Valid Till	15-03-2021
Tender Ref No	KMSCL/EP/T300/341A/2018
Tendered Quantity	100
Supplier Name	M/s Nice Neotech Medical Systems Pvt Ltd
GST No	33AABCN4311M1ZO
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
No.85 Krishna Industrial Estate Vannagaram Mettukuppam Chennai 600 095 Tamil Nadu	Contact Person	K.Premkumar
	Phone	
	Mobile No	9543455323
	Email	sales@niceneotech.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Photo therapy unit single surface LED Type</b> <i>Model &amp; Make : nice 4000 Spot/ Nice Neotech Medical Systems Pvt. Ltd.</i>	30408 Incl.GST :12%	2242.59	32650.59
		<b>30408</b>	<b>2242.59</b>	<b>32650.59</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year	
<b>Photo therapy unit single surface LED Type</b>				
Labour	1,500.00	1,500.00	1,500.00	1,500.00
Comprehensive	2,500.00	2,500.00	2,500.00	2,500.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Photo therapy unit single surface LED Type**

1. Phototherapy should be based on advanced LED technology
2. In LED technology the irradiance should cover the entire treatment area of at least (50 x 25) cms
3. LED Technology.
4. Irradiance level should be 35 microwatts/cm<sup>2</sup> at 35 cms or better distance.
5. Lamp source should be continuous tiltable to 45 degrees angle to cover the entire treatment area.
6. System should be height adjustable at least(110-160) cms from floor level with built-in non – resettable timer
7. The unit should have 3 castor wheels with minimum 2 brakes
8. Should work with input 200 to 240Vac 50 Hz supply.
9. Should supplied with 50 numbers eye masks along with the machine
10. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid.