



Running Contract Details	
Equipment Name	Photo therapy unit double surface LED Type
Running Contract Valid Till	15-03-2021
Tender Ref No	KMSCL/EP/T300/341B/2018
Tendered Quantity	100
Supplier Name	M/s Phoenix Medical Systems Pvt Ltd
GST No	33AAACP1905E1Z2
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
DP 42 SIDCO Industrial Estate Thirumudivakkam Chennai 600 132	Contact Person	Akhil M P
	Phone	
	Mobile No	9388224376
	Email	sales@pmsind.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Photo therapy unit double surface LED Type <i>Model & Make : Brilliance pro with Brilliance Bottom / Phoenix Medical Systems Pvt Ltd</i>	54992 Incl.GST :12%	4055.66	59047.66
		54992	4055.66	59047.66

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 rd Year	4 th Year	5 th Year
Photo therapy unit double surface LED Type			
Labour	1,200.00	1,250.00	1,300.00
Comprehensive	2,000.00	2,100.00	2,200.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Photo therapy unit double surface LED Type

1. Phototherapy should be based on advanced LED technology
2. In LED technology the irradiance should cover the entire treatment area of at least (50 x 25) cms.
3. Should have a height adjustable mechanism, treatment distance to the range of 25 to 45 cms should be possible.
4. Irradiance level should be 35 microwatts/cm² at 35 cms or better distance.
5. Trolley should be mounted in 4 castor wheels with minimum 2 brakes and 3 castor wheels with minimum 2 brakes for upper surface unit
6. The irradiance should be measured and reported to the user institution at the time of installation and thereafter during every subsequent warranty/CMC/AMC visit and all breakdown visits.
7. Lamp source should be continuous tiltable to 45 degrees angle to cover the entire treatment area.
8. System should be height adjustable at least(110-160) cms from floor level with built-in non – resettable timer
9. Baby bed should be transparent with up / down tiltable facility
10. Should work with input 200 to 240Vac 50 Hz supply.
11. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission /FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL.Copy of the certificate/ test report shall be produced along with the technical bid.
12. The measurement of irradiance should be demonstrated to the technical committee during product demonstration
13. Should be supplied with 50 nos Eye masks along with the machine
- ?