



Running Contract Details	
Equipment Name	Power Pack for Electrophoresis System
Running Contract Valid Till	24-04-2021
Tender Ref No	KMSCL/EP/T307/1325/2019(R)
Tendered Quantity	20
Supplier Name	M/s Crescent Lab Equipments
GST No	32AAFFC9887N1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
BMRA-136 B Theckineth B.K Menon Road Edappally Cochin -682024 Kerala	Contact Person	
	Phone	
	Mobile No	
	Email	crescentlab07@gmail.com, info@crescentlab.in,

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Power Pack for Electrophoresis System <i>Model & Make : 1645050/ BIORAD</i>	104425.28 Incl.GST :18%	7309.77	111735.05
		104425.28	7309.77	111735.05

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Power Pack for Electrophoresis System				
Labour	15,000.00	16,500.00		18,150.00
Comprehensive	19,000.00	20,900.00		22,990.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Power Pack for Electrophoresis System

1. Should have power supply 5 – 200 V in 1 V step, 200 W maximum
2. Should have power supply 5 – 200 V in 1 V step, 200 W maximum, 400 mA
3. Output terminals: 4 recessed sets in parallel
4. Timer: 0.999 minutes.
5. Safety features: no load detection, ground leak detection, sudden load change detection, overload/short circuit detection, auto power up after power failure.
6. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid