



Running Contract Details	
Equipment Name	Pulse Oxymeter Model A
Running Contract Valid Till	12-06-2021
Tender Ref No	KMSCL/EP/T322/105A/2019
Tendered Quantity	1500
Supplier Name	M/s BPL Medical Technologies Pvt Ltd
GST No	32AAF3158E1Z4
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
36/227A & 227A-1 MILMA-DAIRY METHANAM ROAD KOONAMTHAI EDAPPALLY COCHIN - 24	Contact Person	BIJU KANADEN
	Phone	0484 2558418
	Mobile No	9846294499
	Email	biju.kanaden@bpl.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Pulse Oxymeter Model A <i>Model & Make : OXY View / BPL Medical Technologies</i>	30240 Incl.GST :12%	2230.2	32470.2
2	Adult reusable Spo2 Probe	3808 Incl.GST :12%	280.84	4088.84
3	One forehead disposable SPO2 sensor	4368 Incl.GST :12%	322.14	4690.14
		38416	2833.18	41249.18

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 rd Year	4 th Year	5 th Year
Pulse Oxymeter Model A			
Labour	810.00	840.00	860.00
Comprehensive	1,600.00	1,700.00	1,900.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Pulse Oximeter Model A

1. Should have plethys-mographic wave form with numeric display for SPO2 and Heart rate on LCD/TFT display screen.
2. Should have a SPO2 range of 0 to 100%.
3. Should have SPO2 accuracy of $\pm 2\%$.
4. Should provide bar graph for pulse strength.
5. Audio and visual alarm for both upper and lower SPO2, Heart rate.
6. Should quote rate separately for Reusable Adult Probe.

The rates offered will be taken for evaluation. (Rate will be fixed for 3 years)

7. Beep sound and alarm sound should have separate volume control
8. Should have a minimum of 4 hours back-up time.
9. Should be a portable, light weight and desktop model with adult, pediatric and Neonatal modes.
10. Should work with input 200 to 240Vac 50 Hz supply.
11. Should have trend data of at least 24 hrs.
12. Should provide with adult reusable finger probe with technology from standard reputed companies like Massimo, Nellcore or equivalent
13. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid
14. Rate of one forehead disposable SPO2 sensor should be quoted separately (if available) which will not be taken for evaluation.