



Running Contract Details	
Equipment Name	Q switched ND Yag laser
Running Contract Valid Till	21-10-2020
Tender Ref No	KMSCL/EP/T267/1158/2018(R)
Tendered Quantity	10
Supplier Name	M/s Aakaar Medical Technologies Pvt Ltd
GST No	27AALCA7587Q1ZK
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
A -801 Heritage Plaza telli galli cross road Andheri east Mumbai - 400069	Contact Person	Mr. Anil Kulkarni
	Phone	02226823949
	Mobile No	9324253669
	Email	anilk@coherentindia.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Q switched ND Yag laser <i>Model & Make : 1064 QCCL / Beijing Honkon Technologies Co.Ltd</i>	1638000 Incl.GST :12%	120802.5	1758802.5
		1638000	120802.5	1758802.5

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Q switched ND Yag laser							
Labour	75,000.00	78,750.00	82,687.50	86,821.88	91,162.96	95,721.11	1,00,507.17
Comprehensive	1,50,000.00	1,57,500.00	1,65,375.00	1,73,643.75	1,82,325.93	1,91,442.23	2,01,014.34

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Q switched ND Yag laser

WAVE LENGTH - 1064/532 NM

OUTPUT ENERGY - SINGLE PULSE-1064NM

DOUBLE PULSE-1064 NM

ENERGY DENSITY(max) - 1064 NM:15.0J/cm²:532nm:8j/cm²

PULSE DURATION - 6 ns

REPETITION RATE - 1-10 Hz adjustable

SPOT SIZE - 2-10 MM ADJUSTABLE

LIGHT GUIDE BEAM - 650 nm RED LASER

COOLING SYSTEM - CLOSED CYCLE WATER TO AIR HEAT EXCHANGER

LASER DELIVERY - 7 JOINTED ARTICULATED ARM

CONTROLLING SYSTEM - 10.4 INCH TFT COLOR TOUCH SCREEN

POWER SUPPLY - 220 VAC,50/60 Hz

1. The laser system to be provided with three additional doctor protective goggles.
2. Should be supplied with UPS of required capacity along with machine for 30 minutes backup.
3. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission/FDA(US)/STQC CB certificate /STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate /test report shall be produced along with the technical bid