

Running Contract Details	
Equipment Name	Refrigerated Cooler for Media Storage
Running Contract Valid Till	24-04-2021
Tender Ref No	KMSCL/EP/T307/1335/2019(R)
Tendered Quantity	100
Supplier Name	M/s Labline Instruments
GST No	32AABFL5818N1Z6
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
P B No. 36 Eroor Road Trippunithura-682301	Contact Person	Bini
	Phone	04842776582
	Mobile No	9020304422
	Email	lablinecochin@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Refrigerated Cooler for Media Storage <i>Model & Make : LMR 10 / Labline Instruments, Cochin</i>	64192 Incl.GST :18%	4493.44	68685.44
		64192	4493.44	68685.44

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Refrigerated Cooler for Media Storage				
Labour	1,100.00	1,100.00	1,100.00	
Comprehensive	4,000.00	4,000.00	5,000.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Refrigerated Cooler for Media Storage

1. Should have upright type unit.
2. Should have a storage Capacity of 300 liters.
3. Should have adjustable temperature control range from +2 to +6 degree Celsius.
4. Should have a hermetically sealed type Compressor.
5. Refrigerant System-
 - a) CFC Free Eco-friendly refrigerant.
 - b) The system should have positive, forced, air circulation to maintain temperature uniformity at all shelf levels, with quick recovery.
 - c) The system should have sensors for activating automatic defrost cycle to minimize the frost build up.
6. Should have a temperature accuracy of +/-0.1 degree.
7. The Inner Cabinet should be made of plastic.
8. The Outer Cabinet made up powder coated corrosion resistant sheet.
9. The Door system should have lockable double glass doors for better safety, (self closing hinge doors)
10. External ambient temperature:- should perform in ambient temperature up to +45 degree Celsius.
11. System should have large and clear digital displays for the set/run parameters.
12. Alarms
 - a. System should have audible/visual warnings for over-temperature, under temperature.
13. Should have Clear product visibility with dual glasses.
14. Should have adjustable plastic coated wire mesh shelves.
15. Number of shelves- 4 Numbers
16. Power input :- 220-230V AC, 50Hz with power cord & plug
17. Should provide suitable voltage stabilizer.

18. Should have FDA/ CE/ UL or BIS approved product.
19. The unit should have fitted with castor wheels.
20. Should have No frost, fan-assisted cooling circulation system.
21. Auto defrost with self-evaporating drip tray.
22. Should have Digital temperature controller display.
23. Should have Full length illumination.