

Running Contract Details	
Equipment Name	Rotary Vacuum Evaporator
Running Contract Valid Till	28-11-2020
Tender Ref No	KMSCL/EP/T280/1227/2018
Tendered Quantity	16
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Rotary Vacuum Evaporator <i>Model & Make : LRE-701A / LABOCON SCIENTIFIC LIMITED,UK</i>	530292 Incl.GST :18%	37120.44	567412.44
		530292	37120.44	567412.44

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Rotary Vacuum Evaporator							
Labour	2,000.00	2,000.00	2,000.00	2,000.00	2,000.00	2,000.00	2,000.00
Comprehensive	4,500.00	4,500.00	4,500.00	4,500.00	4,500.00	4,500.00	4,500.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Rotary Vacuum Evaporator

1. Rotary vacuum evaporator with chiller and vacuum pump
2. Rotation speed: 20-270 RPM (Adjustable)
3. Evaporation flask size: 50 ml to 3000 ml
4. Should accompany with one pair of each size evaporation flask
5. Automatic motorized Lift
6. LED/LCD digital temperature display and control .
7. Temperature accuracy for water bath ± 10 c
8. Over heat protection bath-cut off at 5oc above set point
9. Volume of water bath min 4 liters
10. Glass joint: 29/32(evaporating)
11. Vacuum Seal :PTFE
12. Should have CE issued by a notified body registered in European Commission
13. Necessary tubing to be supplied.
14. IQ/OQ/PQ after satisfactory installation of the Equipment
15. Necessary training should be provided after installation.
16. The interior of the water bath shall be PTFE Coated