



Running Contract Details	
Equipment Name	Rotating atherectomy System (Rotablator)
Running Contract Valid Till	22-06-2020
Tender Ref No	KMSCL/EP/T267/1164/2018(R)
Tendered Quantity	20
Supplier Name	M/s Boston Scientific India Pvt Ltd (VSN MEDICAL CARE PVT LTD)
GST No	32AAECV7124K1ZA
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Rajeev Gandhi Lane Medical College P O Thiruvananthapuram	Contact Person	Mr Partho Banerjee / Mr. Manoj Rath
	Phone	0124-6260500
	Mobile No	9920781992
	Email	hegaindia@bsci.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Rotating atherectomy System (Rotablator) <i>Model & Make : RC 5000 / Boston Scientific India Pvt Ltd</i>	3360000 Incl.GST :12%	247800	3607800
		3360000	247800	3607800

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Rotating atherectomy System (Rotablator)							
Labour	0.00	0.00	0.00	0.00	0.00	0.00	0.00
Comprehensive	1,45,995.00	1,53,295.00	1,60,959.00	1,69,007.00	1,77,458.00	0.00	0.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Rotating atherectomy System (Rotablator)

1. Percutaneous rotational coronary angioplasty with adjunctive balloon angioplasty.
2. To perform Highly Calcific Lesions and diffused disease.
3. 220-240 V A.C Voltage with the frequency of 50-60 Hz.
4. Multiple Rotalink Catheters of varying sizes and with a single Rotalink Advancer for easier more effective debulking.
5. Rotational speed of 0 - 2,50,000 rpm.
6. Console compatibility with DynaGlide system.
7. Diamond coated elliptical burr with sizes 1.25 mm, 1.5 mm, 1.75mm, 2.00mm and 2.25mm, 2.38mm, and 2.5 mm and with Catheter Length of 135 cm in length.
8. Compatibility with 0.014(tip)/.009” Rota wire with 325 cm in length and Floppy, Standard and Extra Support System.
9. Approval International standard agency US FDA or Equivalent.