



Running Contract Details	
Equipment Name	Savory gillard oesophageal dilator with guide
Running Contract Valid Till	14-02-2022
Tender Ref No	KMSCL/EP/T346/1467/2019(R)
Tendered Quantity	20
Supplier Name	M/s Biotrace Impex
GST No	32AAJFB2273P1Z9
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
TC 55/1927(2) Viswambaran Road Pappanamcode P.O Thiruvananthapuram -695018 Kerala	Contact Person	Sreeraj.S
	Phone	
	Mobile No	9446391124
	Email	biotraceimpex@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Savory gillard oesophageal dilator with guide <i>Model & Make : SGD , Cook Medical</i>	190400 Incl.GST :12%	14042	204442
		190400	14042	204442

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Savary gillard oesophageal dilator with guide

1. Dilator length 65-75 cm made of polyvinyl, should have a tapering end
2. Radio opaque band at widest part for location. Should be reusable
3. Set of 10 dilators from 5mm diameter to 14/15mm (5,6,7,8,9,10,11,12,13/12.8,14/15mm)
4. Savary Gillard type guide wire about 200cm long
5. Should provide box to keep dilators and guide wires
6. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid