



Running Contract Details	
Equipment Name	Scanning Laser
Running Contract Valid Till	21-02-2021
Tender Ref No	KMSCL/EP/T296/687/2018(R)
Tendered Quantity	40
Supplier Name	M/s Bharat Medical System
GST No	33AEXPJ4035M1Z2
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
M-1074(New#15) - 7th Main Road Thrivalluvar Nagar Thiruvanniyur Chennai-600041	Contact Person	M Balasubramanian
	Phone	044-24516904/539
	Mobile No	9444052532
	Email	bmsmadras@gmail.com, bmschennai@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Scanning Laser <i>Model & Make : Digi Laser-Pro / Bharat Medical System</i>	363328 Incl.GST :12%	26795.44	390123.44
		363328	26795.44	390123.44

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Scanning Laser							
Labour	2,200.00	2,200.00	2,200.00	2,200.00	2,500.00	2,500.00	2,500.00
Comprehensive	8,600.00	8,600.00	8,600.00	9,000.00	9,000.00	9,000.00	9,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Scanning Laser

1. Should have Visible Red Laser in b/w 620 NM to 670 NM Wave Length
2. Should have Infrared Laser in b/w 800 NM to 850 NM
3. Should have scanning mode with power output of 1.6 W
4. Should have continuous pulsed mode with Duty cycle from 10 - 100 % with a frequency of 1 – 200 Hz.
5. Must be Computerized LCD screen with controls with pre programmed protocols and free memories
6. Should have Automatic electronic sensor
7. Should be mounted on wheels for easy mobility
8. To be supplied with protective goggles and other accessories-Mains cable, Inactive electrode, Scanner.
9. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid
10. Should have a probe with 1200 mW output