



Running Contract Details	
Equipment Name	Short wave diathermy
Running Contract Valid Till	17-01-2021
Tender Ref No	KMSCL/EP/T296/109/2018(R)
Tendered Quantity	50
Supplier Name	M/s Bharat Medical System
GST No	33AEXPJ4035M1Z2
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
M-1074(New#15) - 7th Main Road Thrivalluvar Nagar Thiruvanniyur Chennai-600041	Contact Person	M Balasubramanian
	Phone	044-24516904/539
	Mobile No	9444052532
	Email	bmsmadras@gmail.com, bmschennai@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Short wave diathermy <i>Model & Make : shortwave Diathermy-500W / BHARAT MEDICAL SYSTEMS</i>	175504 Incl.GST :12%	12943.42	188447.42
		175504	12943.42	188447.42

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Short wave diathermy				
Labour	1,500.00	1,500.00	1,500.00	
Comprehensive	3,200.00	3,200.00	3,200.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Short wave diathermy

1. Should be microprocessor controlled solid state technology.
2. Should have continuous and pulsed output.
3. Should have an output of up to 400W in continuous mode.
4. Should have an output of up to 1000W in pulsed mode.
5. Should have a working frequency of 27.1MHz.
6. Should have a pulse width from 20 to 400 micro sec (pulse mode).
7. Should have inductive and capacitive applicator.
8. Should have a selectable treatment time from 1 to 30 minutes.
9. Should be supplied with all standard accessories including disc electrodes, rubber electrodes, mono-polar inductive electrodes, high frequency cables etc.
10. Should work with input 200 to 240 Vac 50Hz supply
11. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.