

Running Contract Details	
Equipment Name	Stainless Steel Patient Cot
Running Contract Valid Till	13-03-2021
Tender Ref No	KMSCL/EP/T311/1220/2019(R)
Tendered Quantity	100
Supplier Name	M/s Molecule Water Technology
GST No	32AOSPK7715G2ZQ
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
Sivalayam Govindamuttom Kayamkulam 690527 Kerala 690527	Contact Person	Prasad.k
	Phone	04792440079
	Mobile No	9400545860
	Email	shkylm@yahoo.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Stainless Steel Patient Cot</b> <i>Model &amp; Make : Molecule Water Technology</i>	17405 Incl.GST :18%	1218.35	18623.35
		<b>17405</b>	<b>1218.35</b>	<b>18623.35</b>

### Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Stainless Steel Patient Cot**

- 1) Should have a stainless steel top grade 304.
  - 2) Should have "Overall dimension shall be 2000mm x 700mm x 800mm (LxWxH) tolerance  $\pm 20$ mm.
  - 3) Should have proper reinforcement for the top.
  - 4) Should be suitable for emergency minor procedures.
- Note: Material test certificate for the proof of usage of SS304 shall be furnished along with invoice and the manufacturer shall also submit undertaking that the tested material is used for the construction of the equipments supplied.