

Running Contract Details	
Equipment Name	Stereo Microscope
Running Contract Valid Till	12-02-2021
Tender Ref No	KMSCL/EP/T299/1337/2018
Tendered Quantity	20
Supplier Name	M/s Radical Scientific Equipments Pvt Ltd
GST No	06AACCR8985N1ZI
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
9th Milestone Ambala-Jagadhri Road Khudda Kalan Ambala Cantt-133104 (Haryana)	Contact Person	Mr. Vikram Singh Chauhan
	Phone	
	Mobile No	89030997827
	Email	tenders@radicalscientific.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Stereo Microscope <i>Model & Make : RSMr-3 / RADICAL</i>	49560 Incl.GST :18%	3469.2	53029.2
2	Compatible adapter for attaching a Nikon DSLR camera	15340 Incl.GST :18%	1073.8	16413.8
		64900	4543	69443

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Stereo Microscope				
Labour	2,100.00	2,100.00	2,100.00	
Comprehensi ve	4,200.00	4,200.00	4,200.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Stereo Microscope

1. The microscope should have capability of giving a laterally erect plan corrected image in a wide field.
2. Should have both transmitted and reflected ring illumination.
3. Should have wide field plan corrected 10X eyepiece of minimum 22mm FOV with dioptr correction.
4. Should have zoom magnification range of up to 50X.
5. Should have an trinocular tube for attaching a camera
6. Price of a compatible adapter for attaching a Nikon DSLR camera should be quoted separately.
7. Should provide dust cover.
8. Should provide two preventive maintenance visits per year and all spares including lamps shall be replaced free of cost during warranty and CMC period.
9. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid