

Running Contract Details	
Equipment Name	STERILIZERâ€™VARIOUS TYPES
Running Contract Valid Till	20-02-2021
Tender Ref No	KMSCL/EP/T307/115/2019(R)
Tendered Quantity	100
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	0% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Steriliser large <i>Model & Make : PE-03 / Paras Enterprises</i>	7552 Incl.GST :18%	528.64	8080.64
2	Sterilizer medium <i>Model & Make : PE-02 / Paras Enterprises</i>	7198 Incl.GST :18%	503.86	7701.86
3	Sterilizer small <i>Model & Make : PE-01 / Paras Enterprises</i>	5133 Incl.GST :18%	359.31	5492.31
		19883	1391.81	21274.81

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.

3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Steriliser large

1. Should have 510mm length, 200mm breadth, 150mm height and 2KW heater coil.
2. Should be made up of seamless stainless steel sheet of grade 304.
3. Should have indicators and power cord.
4. Should operate on 220/230Vac, 50Hz single phase supply.
5. Should have ISI mark.

Equipment :Sterilizer medium

1. Should have 430mm length, 200mm breadth, 150mm height and 1.5KW heater coil.
2. Should be made up of seamless stainless steel sheet of grade 304.
3. Should have indicators and power cord.
4. Should operate on 220/230Vac, 50Hz single phase supply.
5. Should have ISI mark.

Equipment :Sterilizer small

1. Should have 300mm length, 150mm breadth, 125mm height and 1KW heater coil.
2. Should be made up of seamless stainless steel sheet of grade 304.
3. Should have indicators and power cord.
4. Should operate on 220/230Vac, 50Hz single phase supply.
5. Should have ISI mark.