



Running Contract Details	
Equipment Name	Streak Retinoscope
Running Contract Valid Till	26-09-2020
Tender Ref No	KMSCL/EP/T269/118/2018
Tendered Quantity	250
Supplier Name	M/s Inskon Medical Solutions Private Limited
GST No	32AACCI7233A1Z8
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
98C Manthrakkal Junction Thaikkattukara Post Aluva Kochi- 683106	Contact Person	ajesh
	Phone	
	Mobile No	9744160009
	Email	sales@inscon.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Streak Retinoscope</b> <i>Model &amp; Make : 18342 VC / Welch Allyn</i>	28672 Incl.GST :12%	2114.56	30786.56
		<b>28672</b>	<b>2114.56</b>	<b>30786.56</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Streak Retinoscope**

1. Should have an external focusing sleeve which is easy to grip and manipulate.
2. Should have crossed-linear polarizing filter.
3. Should allow one-hand operation for streak focus and 360° streak rotation.
4. Should be interchangeable to plane mirror and concave mirror mode by sleeve movement
5. Should have halogen / LED light source.
6. Should have 100% dust proof housing and multi-coated optics.
7. Should have brow rest for spectacle wearer.
8. Should be rechargeable battery with Charger.
9. Should have a carrying case.
10. If halogen lamp is used, then the following additional accessories should be supplied.
  - a. Bulb – 2 nos.