



Running Contract Details	
Equipment Name	Stretcher Trolley with belt
Running Contract Valid Till	23-10-2020
Tender Ref No	KMSCL/EP/T277/208E/2018
Tendered Quantity	5000
Supplier Name	M/s Medimek Industries
GST No	27ADBPC1957E1ZY
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	0% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
B-50 Additional MIDC Anand Nagar Ambernath(East) - 451506 Thane Maharashtra	Contact Person	Jobin Augustin
	Phone	
	Mobile No	8281815893
	Email	medimek@ymail.com, medimek@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Stretcher Trolley with belt <i>Model & Make : Mi-4081/Medimek Industries</i>	28546.56 Incl.GST :18%	1998.26	30544.82
		28546.56	1998.26	30544.82

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Stretcher Trolley with belt

1. Size : 2000mm (L) x 660mm (W) x 830mm (H)
2. Trolley Main frame : 32mm OD x 1.6 mm SS 304 grade
3. Trolley Supporting frame : 25mm OD x 1.6 mm SS304 grade
4. 4 castor wheels with 200 mm dia.
5. Diagonal Dual lock castors
6. Stretcher top frame 25mm OD x 1.6 mm SS304 grade
7. Sheet should be made of SS 304 grade of 1.2mm thickness
8. Provision for fixing IV pole.
9. Max. Load of 250 Kg
10. Provision for carrying oxygen cylinder.
11. The stretcher shall be detachable from the trolley
12. Mattress shall be provided
13. Patient belt two numbers shall be provided.