



Running Contract Details	
Equipment Name	Strobe Video Rhino Laryngoscope with monitor camera and Light Source
Running Contract Valid Till	07-01-2021
Tender Ref No	KMSCL/EP/T293/1296/2018
Tendered Quantity	16
Supplier Name	M/s Praxis Health Technologies
GST No	32AAPFP1540D1ZM
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Puthiyapalam Tali Kozhikode Chalappuram P.O	Contact Person	Anoop M
	Phone	4952700425
	Mobile No	9747275757
	Email	praxis.health2013@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Strobe Video Rhino Laryngoscope with monitor camera and Light Source</b> <i>Model &amp; Make : CAM 41 HD(Camera), strobo 21 LED(Strobe) / ATMOS MedizinTechnik GmbH &amp; Co. KG</i>	2005610.64 Incl.GST :0%	165663.44	2171274.08
		<b>2005610.64</b>	<b>165663.44</b>	<b>2171274.08</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Strobe Video Rhino Laryngoscope with monitor camera and Light Source</b>							
Labour	67,700.38	76,162.93	84,625.48	93,088.03	1,01,550.57	1,10,013.19	11,84,475.66
Comprehensive	1,01,550.57	1,10,013.19	1,18,475.67	1,26,938.21	1,35,400.76	1,43,863.39	1,52,325.86

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Strobe Video Rhino Laryngoscope with monitor camera and Light Source**

##### A. MONITOR

- a) Power input : 100 W
- b) Power supply : 100 – 240 V AC
- c) Interfaces :
  - i. video interfaces : DVI – D (in/out)
  - ii. audio interface
  - iii. foot switch port : five pin socket for foot switch
  - iv. printer port : USB
  - v. Foot switch for stroboscopy
- d) Light source : LED colour temperature of 6000 – 7000 K and  
not less than 30,000 hours life span.
- e) Image format : JPG
- f) Storage : USB / memory card/ Hard disk
- g) Monitor :
  - i. screen size :14-15”
  - ii. resolution : Good
  - iii. Contrast : 700 – 800 : 1
- h) Loud speaker : 2W power
- i) Stroboscopy kit :
  - i. USB footswitch with integrated activation key for stroboscopy function

ii. Microphone set including microphone fitting, microphone holder, connector for microphone membrane.

j) Camera head : single chip camera.

#### B. STROBO VIDEO RHINO- LARYNGOSCOPE

a) Chip on tip camera

b) Deflection up / down 140 °/140 °.

c) Direction of view 0°.

d) Angle of view 85°.

e) Distal end outer diameter 3.7 mm.

f) Working length 30 cm.

#### C. COMPATIBLE RIGID ENDOSCOPE:

a) Laryngoscope

b) Integrated Hopkins rod lateral telescope 70°.

c) Angle of view 50°

d) Diameter 8 mm.

e) Length (16-18) cm.

f) Autoclavable with fibre optic light transmission with handle.

#### D. ACCESSORIES.

a) USB flash drive 32 GB/ Hard disk

b) Suitable UPS.

E. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.