



Running Contract Details	
Equipment Name	Temperature controlled waterbath
Running Contract Valid Till	07-01-2021
Tender Ref No	KMSCL/EP/T280/142C/2018
Tendered Quantity	28
Supplier Name	M/s Kochin Inlab Equipments India Pvt Ltd
GST No	32AADCK3263F1ZX
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Door #28/2391-D Vihaan - Chilavannur Kadavanthra P.O Kochi-682020.	Contact Person	Manoj
	Phone	
	Mobile No	9447790938
	Email	manoj.inlab@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Temperature controlled waterbath</b> <i>Model &amp; Make : BSW-04D / BESTON INDUSTRIES, KOCHI</i>	13393 Incl.GST :18%	937.51	14330.51
		<b>13393</b>	<b>937.51</b>	<b>14330.51</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year	
<b>Temperature controlled waterbath</b>				
Labour	500.00	550.00	600.00	
Comprehensive	1,000.00	1,100.00	1,210.00	

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Temperature controlled waterbath**

1. Should have a double walled construction.
2. The inner chamber and top lid should be made of stainless steel.
3. The space between the two walls should be packed with thick glass wool.
4. Should provide with a microprocessor based variable digital temperature controller with digital display.
5. Working temperature should be from ambient to 800C having an accuracy of +/-10C
6. Should have an approximate inner chamber dimension of 450mm x 300mm x 175mm