



Running Contract Details	
Equipment Name	TENS (Table Top 4 Channel)
Running Contract Valid Till	20-06-2021
Tender Ref No	KMSCL/EP/T319/1316/2019(R)
Tendered Quantity	100
Supplier Name	M/s Bharat Medical System
GST No	33AEXPJ4035M1Z2
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
M-1074(New#15) - 7th Main Road Thrivalluvar Nagar Thiruvanimiyur Chennai-600041	Contact Person	M Balasubramanian
	Phone	044-24516904/539
	Mobile No	9444052532
	Email	bmsmadras@gmail.com, bmschennai@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>TENS (Table Top 4 Channel)</b> <i>Model &amp; Make : Digi Tens Pro / BHARAT MEDICAL SYSTEMS</i>	13888 Incl.GST :12%	1024.24	14912.24
		<b>13888</b>	<b>1024.24</b>	<b>14912.24</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year	
<b>TENS (Table Top 4 Channel)</b>				
Labour	85.00	85.00	85.00	
Comprehensive	255.00	255.00	255.00	

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :TENS (Table Top 4 Channel)**

1. It should be an advanced microprocessor with minimum 4 channel TENS machine.
2. It should be Sophisticated, compact, lightweight with digital display of patient parameters.
3. It should have an automatic intensity control to make it safe to operate.
4. Number of channels- minimum 4 (independent).
5. It should have independent intensity control and display of output current for each channel.
6. It should have digital display of output current of each channel and treatment time.
7. The machine should produce output only after intensity of all channels are set to Zero.
8. Output current -0-60 mA
9. Output voltage 0-100V
10. Frequency 1-200 Hz
11. Pulse width 75-360 sec
12. Treatment time- 5,10,15,20 minutes.
13. Unit should function with 200-240 AC/ 50- 60 Hz input power supply.
14. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL.