

Running Contract Details	
Equipment Name	Therapeutic Double channel Gastroscope with compatible processor
Running Contract Valid Till	21-10-2020
Tender Ref No	KMSCL/EP/T267/1067/2018(R)
Tendered Quantity	5
Supplier Name	M/s Cardinal Medical Technologies Private Limited
GST No	33AADCC0058L1ZV
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
5/721 - Sabari Salai Madipakkam Chennai - 600091 Tamil Nadu	Contact Person	Mr. P. Srinivas
	Phone	044-22581678/88
	Mobile No	9791032492
	Email	cardinalmedtech@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Therapeutic Double channel Gastroscope with compatible processor <i>Model & Make : EG_3890 TK , EPK -3000(L) / Pentax Medical</i>	1960000 Incl.GST :12%	144550	2104550
		1960000	144550	2104550

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Therapeutic Double channel Gastroscope with compatible processor							
Labour	21,186.00	21,186.00	23,305.00	23,305.00	25,424.00	25,424.00	25,424.00
Comprehensive	50,847.00	50,847.00	50,847.00	57,627.00	57,627.00	61,017.00	61,017.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Therapeutic Double channel Gastroscope with compatible processor

1. Viewing direction of the scope should be forward (0°)
2. It should have minimum 140° field of view
3. It should have maximum angulation as follows:
 - a. Up: 180° or more
 - b. Down: 90° or more
 - c. Right: 100° or more
 - d. Left: 100° or more
4. Focal range should be 3-100mm approximately
5. Insertion tube diameter should be of maximum of 12.8 mm
6. Distal end diameter should be maximum of 13.5mm
7. Instrument channels should be of 2.5 – 4mm diameter each
8. It should have working length of 1030-1100mm
9. Total length should be of 1300 -1450 mm
10. It should have forward water jet mechanism

II. Video Processor

1. It should be compact and light weight
2. Should have integrated Xenon light source
3. Should have a halogen/ LED lamp as emergency standby
4. Lamp: 100 W or more

5. Video outputs: RGB, S-Video, Composite

6. Should have a provision for brightness adjustment on machine

7. Should have a lamp cooling mechanism

8. Power supply: 120 V-230V, 50/60Hz

III. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid

IV. A monitor with 17" or more should be supplied along with the unit

V. A system cart of suitable size for placing video processor and monitor should be supplied