



Running Contract Details	
Equipment Name	Tilt Table
Running Contract Valid Till	22-02-2021
Tender Ref No	KMSCL/EP/T300/348/2018
Tendered Quantity	25
Supplier Name	M/s Bharat Medical System
GST No	33AEXPJ4035M1Z2
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
M-1074(New#15) - 7th Main Road Thrivalluvar Nagar Thiruvanniyur Chennai-600041	Contact Person	M Balasubramanian
	Phone	044-24516904/539
	Mobile No	9444052532
	Email	bmsmadras@gmail.com, bmschennai@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Tilt Table <i>Model & Make : Tilt Table / BHARAT MEDICAL SYSTEMS</i>	54320 Incl.GST :12%	4006.1	58326.1
		54320	4006.1	58326.1

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Tilt Table				
Labour	485.00	485.00	485.00	
Comprehensive	970.00	970.00	970.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Tilt Table

1. Electrically operated tilt table with height adjustable system
2. It should have a foam padded top, provided with 3 straps to hold the patient.
3. There should be a heavy duty motor which provide silent and smooth lifting action with push button control switches.
4. Range of tilt 0-90 deg.
5. The size of table top 60 cm wide*195cm long*80cm high and fitted in a heavy duty tubular steel frame with four wheels of 100 mm diameter. For easy mobility
6. There should be a utility tray and gripping handle for ADL activity.
7. It should have a oven baked finish
8. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL