



Running Contract Details	
Equipment Name	Tissue Homogenizer Model B
Running Contract Valid Till	13-03-2020
Tender Ref No	KMSCL/EP/T262/1107/2017(R)
Tendered Quantity	10
Supplier Name	M/s Labindia Instruments Pvt Ltd
GST No	33AAACL5140P1ZG
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
B-1 Alsa Regency #165 Eldams Road Alwarpet Chennai- 600018	Contact Person	Mr.Pramod P G
	Phone	022 25986037
	Mobile No	08129106082
	Email	ramananj@labindia.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Tissue Homogenizer Model B <i>Model & Make : 220V Omni Tissue Master 125 with 5 mm Probe / OMNI International</i>	153942.8 Incl.GST :18%	10776	164718.8
		153942.8	10776	164718.8

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Tissue Homogenizer Model B							
Labour	33,040.00	37,170.00	41,300.00	46,020.00	51,330.00	57,230.00	63,130.00
Comprehensive	49,560.00	55,460.00	61,950.00	69,030.00	76,700.00	85,550.00	94,990.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Tissue Homogenizer Model B

1. The equipment should be able to disperse small quantity of biological material into 5 – 10 μ m sized suspension.
2. The device should be compact
3. It should have removable dispersal elements with a rotor stator mechanism.
4. Should be able to process volumes as low as 0.5 ml.
5. The dispersing element should be autoclavable and resistant to all common disinfectants.
6. The unit should be supplied with 2 nos. of dispersing tools.
7. Should work in 220/230 VAC, 50 Hz.
8. The equipment should be supplied with all accessories required for normal operation.
9. Price of any accessories and spare dispersing elements should be quoted separately.